

## Is self monitoring of blood pressure in pregnancy safe and effective?

Hodgkinson, James A; Tucker, Katherine L; Crawford, Carole; Greenfield, Sheila M; Heneghan, Carl; Hinton, Lisa; Khan, Khalid; Locock, Louise; Mackillop, Lucy; McCourt, Christine; Selwood, Mary; McManus, Richard J

DOI:  
[10.1136/bmj.g6616](https://doi.org/10.1136/bmj.g6616)

License:  
None: All rights reserved

*Document Version*  
Publisher's PDF, also known as Version of record

*Citation for published version (Harvard):*  
Hodgkinson, JA, Tucker, KL, Crawford, C, Greenfield, SM, Heneghan, C, Hinton, L, Khan, K, Locock, L, Mackillop, L, McCourt, C, Selwood, M & McManus, RJ 2014, 'Is self monitoring of blood pressure in pregnancy safe and effective?', *BMJ*, vol. 349, g6616. <https://doi.org/10.1136/bmj.g6616>

[Link to publication on Research at Birmingham portal](#)

### **Publisher Rights Statement:**

Article published as: Hodgkinson, James A., et al. "Is self monitoring of blood pressure in pregnancy safe and effective?." *BMJ* 349 (2014): g6616.

Available online: <http://dx.doi.org/10.1136/bmj.g6616>

Checked October 2015

### **General rights**

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

### **Take down policy**

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact [UBIRA@lists.bham.ac.uk](mailto:UBIRA@lists.bham.ac.uk) providing details and we will remove access to the work immediately and investigate.

# ANALYSIS

## Is self monitoring of blood pressure in pregnancy safe and effective?

James A Hodgkinson *research fellow*<sup>1</sup>, Katherine L Tucker *research fellow*<sup>2</sup>, Carole Crawford *research midwife*<sup>2</sup>, Sheila M Greenfield *professor of medical sociology*<sup>1</sup>, Carl Heneghan *professor of evidence based medicine*<sup>2</sup>, Lisa Hinton *senior qualitative researcher*<sup>2</sup>, Khalid Khan *professor of obstetrics*<sup>3</sup>, Louise Locock *associate professor*<sup>2,4</sup>, Lucy Mackillop *obstetric physician*<sup>5</sup>, Christine McCourt *professor of maternal health*<sup>6</sup>, Mary Selwood *midwife*<sup>2</sup>, Richard J McManus *National Institute for Health Research professor*<sup>2</sup>

<sup>1</sup>Primary Care Clinical Sciences, University of Birmingham, Birmingham, UK; <sup>2</sup>Nuffield Department of Primary Care Health Sciences, University of Oxford, Oxford OX3 6GG, UK; <sup>3</sup>Women's Health Research Unit, Centre for Primary Care and Public Health, Blizard Institute, London, UK; <sup>4</sup>NIHR Oxford Biomedical Research Centre, Churchill Hospital, Oxford, UK; <sup>5</sup>Oxford University Hospitals NHS Trust, Women's Centre, John Radcliffe Hospital, Oxford, UK; <sup>6</sup>City University London, London, UK

### Abstract

Guidelines encourage the use of self monitoring of blood pressure in pregnancy, and research suggests that women prefer it. But **Hodgkinson and colleagues** explain that our enthusiasm may run ahead of the evidence and call for more research before it is routinely adopted

Self monitoring of blood pressure is increasingly popular with patients and healthcare professionals. Around a third of people with hypertension self monitor,<sup>1</sup> and measurements are more accurate than readings taken in clinic.<sup>2</sup> Anecdotal reports in the UK suggest that self monitoring in pregnancy is commonplace, although no studies have assessed this. A Canadian pilot survey found that two thirds of women with gestational hypertension were self monitoring.<sup>3</sup> Another small Canadian survey found that 78% of obstetricians used self monitoring in preference to ambulatory measurement to check for white coat hypertension in pregnant women with raised blood pressure.<sup>4</sup>

Blood pressure guidelines recommend home monitoring for pregnant women with chronic hypertension and poorly controlled blood pressure and for women with gestational hypertension—for example, the 2013 American College of Obstetricians and Gynaecologists guidelines<sup>5</sup>—so it is likely that the practice will become more common. The American Heart Association, American Society of Hypertension, and Preventive Cardiovascular Nurses Association joint statement<sup>6</sup> and European Society of Hypertension guidelines have highlighted the importance and potential of self monitoring blood pressure, with the American guidelines describing it as “theoretically ideal for monitoring changes in blood pressure during pregnancy.”<sup>7</sup>

Although home monitoring in pregnancy may have some advantages, there are still many unanswered questions about its use (box 1). We discuss the available evidence on self monitoring and suggest a way forward.

### How many women are affected?

Hypertensive disorders during pregnancy (box 2) are a leading cause of maternal mortality worldwide.<sup>8</sup> They are also associated with fetal growth restriction, low birth weight, preterm delivery, respiratory distress syndrome, and admission to neonatal intensive care.<sup>9</sup>

Substandard care has been identified in 46% of maternal deaths from pre-eclampsia and 65% of fetal deaths associated with pre-eclampsia.<sup>11</sup> In particular, failures to identify and act on known risk factors at booking appointments and to recognise and respond to signs and symptoms from 20 weeks' gestation have been noted.

A tenth of women have raised blood pressure (>140/90 mm Hg), with or without proteinuria, during pregnancy worldwide,<sup>12</sup> and the proportion of women with high blood pressure, and risk factors for high blood pressure, is increasing. Obesity, for example, is associated with a threefold increase in risk of pre-eclampsia,<sup>13 14</sup> and in the US the percentage of women who are obese (body mass index >30) or overweight (> 25) has increased almost 60% in the past 30 years. Similarly, advanced maternal age is associated with around a 50% increase of pre-eclampsia,<sup>15</sup> and the number of pregnant women over 40 in the UK has more than doubled in the past 24 years.<sup>16</sup>

**Box 1: Advantages and disadvantages of home monitoring of blood pressure in pregnancy***Potential advantages*

- Increased accuracy
- Patient friendly
- Potential to free healthcare professional time or reduce clinic visits
- Potential to identify white coat hypertension
- May reduce women's anxiety or medicalisation of care

*Potential disadvantages*

- Few monitors have been validated for use in pregnancy
- Poor understanding of normal blood pressure in pregnancy
- No diagnostic thresholds from home monitoring to identify pre-eclampsia or gestational hypertension
- False reassurance if woman with white coat hypertension subsequently develops true hypertension in pregnancy
- No evidence that earlier detection of high blood pressure through home monitoring will alter outcomes
- No evidence on optimal frequency and timing of home monitoring
- May increase women's anxiety or medicalisation of care

**Box 2: Types of hypertension during pregnancy**

*Chronic hypertension*—Blood pressure >140/90 mm Hg before 20 weeks or being treated at time of referral to maternity services

*Gestational hypertension*—New hypertension presenting after 20 weeks without significant proteinuria

*Pre-eclampsia*—New hypertension presenting after 20 weeks of pregnancy combined with significant proteinuria (protein:creatinine ratio>30 mg/mmol or validated 24 h urine collection containing >300 mg protein)<sup>10</sup>

## Blood pressure in pregnancy

Pre-eclampsia can manifest long before women experience physical symptoms,<sup>17 18</sup> and there is evidence that women in the UK develop the condition between antenatal visits. Of 383 women with confirmed eclampsia, 323 (85%) had been seen by a doctor or midwife in the week before their first convulsion, but at that point 36 (11%) had neither hypertension nor proteinuria, 32 (10%) had proteinuria but no hypertension, and 71 (22%) had hypertension alone.<sup>19</sup>

Current guidelines recommend blood pressure monitoring at routine antenatal visits with “increased frequency” for those at higher risk.<sup>20</sup> Comprehensive systematic reviews have not identified any specific screening test for pre-eclampsia (including those based on demographic characteristics, biomarkers, and ultrasound screening) that has sufficient accuracy or cost effectiveness to introduce into clinical practice. Consequently blood pressure monitoring in clinics remains the mainstay of pre-eclampsia detection in antenatal care.<sup>21</sup> Earlier identification of rising blood pressure in asymptomatic women would allow better targeting of resources at those in need of close monitoring, and self measurement between clinic appointments could facilitate this.

Clinic measurements are more vulnerable to error than out of office monitoring; more measurements can be made at home and white coat hypertension is avoided. The variability of blood pressure is heightened in pregnancy,<sup>22</sup> and debate continues about what constitutes normal blood pressure in pregnancy and how this may change by trimester. Reliance on clinic measurements could lead to unnecessary monitoring or missed opportunities to detect raised blood pressure, though there is no reliable evidence of this being the case. Data from ambulatory monitoring suggest that outcomes in women with white coat hypertension in pregnancy are similar to those with normal blood pressure.<sup>23</sup>

## Will women self monitor?

We know pregnant women are willing to undertake repeated self measurements, comply with monitoring schedules,<sup>10</sup> and are able to accurately record blood pressure data.<sup>24</sup> It does not seem to increase anxiety,<sup>25</sup> even if more complex telemonitoring equipment is used.<sup>26 27</sup> Clinicians promoting self monitoring report being encouraged by women's cooperation, competence, and genuine desire to participate in their healthcare.<sup>28 29</sup> Over 98% of women with hypertension in pregnancy reported liking being involved in their blood pressure management.<sup>3</sup>

Self monitoring of blood pressure in pregnancy is more acceptable to pregnant women than more frequent clinic visits,<sup>30</sup> hospital admission,<sup>27</sup> or ambulatory monitoring.<sup>31 32</sup> Among 78 healthy pregnant women who evaluated both home and ambulatory monitors at 35-37 weeks' gestation, 74 (95%) found self monitoring of blood pressure (using Omron HEM705CP) acceptable compared with 61 (78%) for 24 hour ambulatory monitoring.<sup>33</sup> Home monitoring caused less discomfort and rarely interfered with activities or disturbed sleep.

## Is self monitoring accurate?

Despite the large number of home monitors available, few are validated and deemed accurate for use in pregnancy and pre-eclampsia. Specific validation of monitors in pregnant women is important because several monitors validated for general use have been found to be inaccurate, mostly because of falsely low readings.<sup>34 35</sup> Five monitors have been validated for home use in pregnancy using widely accepted protocols.<sup>36-40</sup> But even validated monitors may not be accurate for all pregnant women—for example, when the accuracy of three commercially available devices was tested on 55 pregnant women with upper arm circumference >35 cm, none was accurate.<sup>41</sup>

## How should self monitoring be done?

In essential hypertension, self monitoring for a minimum of three days and ideally seven days is currently recommended,

although the evidence underlying this is not particularly compelling.<sup>42</sup> Conclusions from ambulatory monitoring in pregnancy suggest that a blood pressure of 135/85 mm Hg best predicts future pregnancy induced hypertension, but the threshold has not been established firmly.<sup>43-44</sup> Clear self monitoring thresholds for hypertension in pregnancy have not been established.<sup>42-45</sup> Few data are available comparing clinic thresholds with self monitored blood pressure in pregnancy, and the studies they are drawn from have important methodological weaknesses.

At a more basic level incomplete understanding of normal blood pressure in pregnancy means that any monitoring in pregnancy is challenging. Blood pressure changes through the trimesters, falling and then rising again; thus, a woman could have a blood pressure that starts at 100/70 mm Hg, falls in mid-trimester to 90/60 mm Hg, and then rises to 135/85 mm Hg (for example) but still be considered within normal limits (<140/90 mm Hg by clinic measurement). It is unclear whether monitoring frequency should change if blood pressure rises notably but remains below the threshold of 140/90 mm Hg.<sup>46</sup> Despite this, at present only the 140/90 mm Hg threshold is accepted for all trimesters.<sup>10</sup> Women with pre-existing hypertension may have more unpredictable blood pressure in pregnancy because of stopping treatment or changing to safer medicines in early pregnancy.

In the absence of clear evidence, there is little guidance on how often blood pressure should be measured. The US guidelines recommend that in women with hypertension before pregnancy “the diagnosis should be confirmed by multiple measurements and may incorporate home or other out-of-office blood pressure readings,” but they do not say what this means practically—for example, how frequently blood pressure should be measured.<sup>47</sup> NICE guidelines on hypertension in pregnancy conclude that research is needed to determine the optimal frequency and timing of measurement as well as on the best way to detect proteinuria in women who have existing hypertension or other known risk factors for pre-eclampsia.<sup>48</sup>

## Does self monitoring affect pregnancy outcomes?

It is not known whether self monitoring will alter outcomes. In the UK, a pilot randomised controlled trial in 80 low risk women included weekly self monitoring combined with a reduced antenatal visit schedule.<sup>49</sup> A larger trial did not go ahead, perhaps because the predicted number of low risk women required to determine a significant effect was 10 000. A future trial may be better to focus on the role of self monitoring in a higher risk group, which would need fewer women because they would be more likely to benefit.<sup>50</sup>

Few data exist on the safety of self monitoring, but there are conflicting reports regarding how well women follow instructions from healthcare professionals.<sup>30-51</sup> In one small study, women who recorded their own blood pressure responded appropriately by contacting healthcare professionals when repeated readings were persistently raised.<sup>52</sup> However, researchers found 10 out of 21 pregnant women in another study<sup>36</sup> had poor understanding of the instructions given about the importance of alerting midwives when their blood pressure was raised above a threshold. Problems included language barriers and personal or work commitments.

## The bottom line

Self monitoring of blood pressure seems to be feasible and acceptable to pregnant women. It might make antenatal care more effective, but we need further research to establish safety and efficacy, the impact on women and health professionals, and how best to use the results. If self monitoring becomes widely used within healthcare systems, such research may be difficult to undertake. We believe that until the evidence base is considerably stronger, further implementation of self monitoring of blood pressure in pregnancy, at least formally by the NHS, should be delayed.

Nevertheless, the trend towards self monitoring is set to increase, and it is important to acknowledge this and respect people's choices. General practitioners and midwives need to be sensitive to this, and could ask women in their care whether they are self monitoring and, if so, whether they would like to share the results or receive any help interpreting them. They should also discuss the uncertainty around how the results are best interpreted.

**Contributors and sources:** All authors are members of the Blood Pressure Monitoring in Pregnancy Research Group (BUMP). JAH and KLT are research fellows in primary care; RJMcM and CH are academic GPs; CC and MS are both midwives and researchers; SMG, LH, and LL are qualitative researchers with expertise in personal experiences of health and illness; KK, LM, and CMcC work in obstetrics and maternal health. Key studies were identified from Medline searches up to March 2014, without limiting by publication date or language and searches of BHS and dabl websites. JAH, KLT, and RJMcM wrote the first draft of the article, and subsequent drafts were edited for critical content by all. RJMcM is guarantor.

**Competing interests:** We have read and understood BMJ policy on declaration of interests and declare the following interests: RJMcM is a member of the NIHR HTA Trials Board and has received blood pressure monitoring equipment for research studies from Omron and Lloyds Pharmacy.

This article forms part of a larger study of blood pressure monitoring in pregnancy funded by the NIHR National School of Primary Care. We acknowledge the support of the NIHR Oxford Collaboration for Leadership in Applied Health Research and Care, and the invaluable contribution of Tricia Carver, Margaret Glogowska, and Ursula Saunders, patient representatives, in discussions relating to this review. We thank Jim Thornton and Lucy Chappell, who reviewed a previous version of this article.

**Provenance and peer review:** Not commissioned; externally peer reviewed.

- 1 Baral-Grant S, Haque MS, Nouwen A, Greenfield SM, McManus RJ. Self-monitoring of blood pressure in hypertension: a UK primary care survey. *Int J Hypertens* 2012;58:2068.
- 2 Hodgkinson J, Mant J, Martin U, Guo B, Hobbs FDR, Deeks JJ, et al. Relative effectiveness of clinic and home blood pressure monitoring compared to ambulatory blood pressure monitoring in the diagnosis of hypertension: a systematic review. *BMJ* 2011;342:d3621
- 3 Magee LA, von Dadelszen P, Chan S, Gafni A, Gruslin A, Helewa M, et al. Women's views of their experiences in the CHIPS (Control of Hypertension in Pregnancy Study) pilot trial. *Hypertens Pregnancy* 2007;26:371-87.
- 4 Dehaeck U, Thurston J, Gibson P, Stephanson K, Ross S. Blood pressure measurement for hypertension in pregnancy. *J Obstet Gynaecol Can* 2010;32:328-34.
- 5 American College of Obstetricians and Gynaecologists. Hypertension, pregnancy-induced. Practice guideline. ACOG, 2013.
- 6 Pickering TG, Miller NH, Ogedegbe G, Krakoff LR, Artinian NT, Goff D. Call to action on use and reimbursement for home blood pressure monitoring: a joint scientific statement from the American Heart Association, American Society of Hypertension, and Preventive Cardiovascular Nurses Association. *Hypertension* 2008;52:10-29.
- 7 Parati G, Stergiou GS, Asmar R, Bilo G, de Leeuw P, Imai Y, et al. European Society of Hypertension guidelines for blood pressure monitoring at home: a summary report of the second international consensus conference on home blood pressure monitoring. *J Hypertens* 2008;26:1505-26.
- 8 Khan KS, Wojdyla D, Say L, Gulmezoglu AM, Van Look PF. WHO analysis of causes of maternal death: a systematic review. *Lancet* 2006;367:1066-74.
- 9 Altman D, Carroli G, Duley L, Farrell B, Moodley J, Neilson J, et al. Do women with pre-eclampsia and their babies benefit from magnesium sulphate? The Magpie Trial: a randomised placebo-controlled trial. *Lancet* 2002;359:1877-90.

- 10 National Institute for Health and Clinical Excellence. Hypertension in pregnancy. NICE clinical guideline CG107. NICE, 2010.
- 11 Maternal and Child Health Research Consortium. Confidential enquiry into stillbirths and deaths in infancy: 8th annual report, 1 January-31 December 1999. Maternal and Child Health Research Consortium, 2001.
- 12 Duley L. The global impact of pre-eclampsia and eclampsia. *Semin Perinatol* 2009;33:130-7.
- 13 Bodnar LM, Ness RB, Markovic N, Roberts JM. The risk of preeclampsia rises with increasing prepregnancy body mass index. *Ann Epidemiol* 2005;15:475-82.
- 14 Wang Y, Beydoun MA, Liang L, Caballero B, Kumanyika SK. Will all Americans become overweight or obese? estimating the progression and cost of the US obesity epidemic. *Obesity* 2008;16:2323-30.
- 15 Lamminpää R, Vehviläinen-Julkunen K, Gissler M, Heinonen S. Preeclampsia complicated by advanced maternal age: a registry-based study on primiparous women in Finland 1997-2008. *BMC Pregnancy Childbirth* 2012;12:47.
- 16 Office for National Statistics. Statistical bulletin: conceptions in England and Wales 2012. ONS, 2014.
- 17 Khong TY, De Wolf F, Robertson WB, Brosens I. Inadequate maternal vascular response to placentation in pregnancies complicated by pre-eclampsia and by small-for-gestational age infants. *Br J Obstet Gynaecol* 1986;93:1049-59.
- 18 Redman CW, Sargent IL. Latest advances in understanding preeclampsia. *Science* 2005;308:1592-4.
- 19 Douglas KA, Redman CW. Eclampsia in the United Kingdom. *BMJ* 1994;309:1395-1400.
- 20 National Institute for Health and Care Excellence. Antenatal care. NICE clinical guideline CG62. NICE, 2008.
- 21 Meads CA, Cnossen JS, Meher S, Juarez-Garcia A, ter Riet G, Duley L, et al. Methods of prediction and prevention of pre-eclampsia: systematic reviews of accuracy and effectiveness literature with economic modelling. *Health Technol Assess* 2008;12.
- 22 Ayala DE, Hermida RC, Mojón A, Fernández JR, Silva I, Uciada R, et al. Blood pressure variability during gestation in healthy and complicated pregnancies. *Hypertension* 1997;30:611-8.
- 23 Bellomo G, Narducci PL, Swifts F, Pastorelli G, Stangoni G, Angeli G, et al. Prognostic value of 24-hour blood pressure in pregnancy. *JAMA* 1999;282:1447-52.
- 24 Kitagawa M, Akiyama Y, Omi H, Sago H, Natori M. Development and clinical application of a telemedicine support system in the field of perinatal patient management. *J Obstet Gynaecol Res* 2000;26:427-34.
- 25 Rayburn WF, Zuspan FP, Piehl EJ. Self-monitoring of blood pressure during pregnancy. *Am J Obstet Gynecol*. 1984;148:159-62.
- 26 Swindells HE, Mooney P, Cartwright W, Dalton KJ. Blood pressure telemetry from home. *Midwife Health Visitor Community Nurse* 1990;26:88-90.
- 27 Cartwright W, Dalton KJ, Swindells H, Rushant S, Mooney P. Home measurement of pregnancy hypertension. *Professional Care Mother Child* 1993;3:8-9.
- 28 Rayburn WF, Zuspan FP, Piehl EJ. Self-monitoring of blood pressure during pregnancy. *Am J Obstet Gynecol* 1984;148:159-62.
- 29 Rayburn WF, Piehl EJ, Compton AA. Clinical significance of home blood pressure monitoring during pregnancy. *Clin Exp Hypertens Part B* 1985;4:63-73.
- 30 Naef RW 3rd, Perry KG Jr, Magann EF, McLaughlin BN, Chauhan SP, Morrison JC. Home blood pressure monitoring for pregnant patients with hypertension. *J Perinatol* 1998;18:226-9.
- 31 Cartwright W, Dalton KJ, Swindells H, Rushant S, Mooney P. Objective measurement of anxiety in hypertensive pregnant women managed in hospital and in the community. *Br J Obstet Gynaecol* 1992;99:182-5.
- 32 Denolle T, Weber JL, Calvez C, Getin Y, Daniel JC, Lurton O, et al. Diagnosis of white coat hypertension in pregnant women with teletransmitted home blood pressure. *Hypertens Pregnancy* 2008;27:305-13.
- 33 Taylor RS, Freeman L, North RA. Evaluation of ambulatory and self-initiated blood pressure monitors by pregnant and postpartum women. *Hypertens Pregnancy* 2001;20:25-33.
- 34 Reinders A, Cuckson AC, Jones CR, Poet R, O'Sullivan G, Shennan AH. Validation of the Welch Allyn 'Vital Signs' blood pressure measurement device in pregnancy and pre-eclampsia. *BJOG* 2003;110:134-8.
- 35 Natarajan P, Shennan AH, Penny J, Halligan AW, de Swiet M, Anthony J. Comparison of auscultatory and oscillometric automated blood pressure monitors in the setting of preeclampsia. *Am J Obstet Gynecol* 1999;181:1203-10.
- 36 Chung Y, de Greeff A, Shennan A. Validation and compliance of a home monitoring device in pregnancy: Microlife WatchBP Home. *Hypertens Pregnancy* 2009;28:348-359.
- 37 De Greeff A, Reggiori F, Anthony J, Shennan A. The Microlife 3AC1: an accurate blood pressure measurement device in pregnancy and pre-eclampsia. *J Hypertens* 2006;24(suppl 4):S279.
- 38 Reinders A, Cuckson AC, Lee JTM, Shennan AH. An accurate automated blood pressure device for use in pregnancy and pre-eclampsia: the Microlife 3BTO-A. *BJOG* 2005;112:1-6.
- 39 De Greeff A, Beg Z, Gangji Z, Dorney E, Shennan AH. Accuracy of inflationary versus deflationary oscillometry in pregnancy and preeclampsia; Omron MIT versus Omron M7. *Blood Pressure Monit* 2009;14:37-40.
- 40 Golar M, Benedict A, Jones C, Randhawa M, Poston L, Shennan AH. Inflationary oscillometry provides accurate measurement of blood pressure in pre-eclampsia *Br J Obstet Gynaecol* 2002;109:1143-7.
- 41 Schwartz WJ 3rd, Rayburn WF, Turnbull GL, Christensen HD. Blood pressure monitoring during pregnancy. Accuracy of portable devices designed for obese patients. *J Reprod Med* 1996;41:581-5.
- 42 NICE. Hypertension: clinical management of primary hypertension in adults. NICE guideline CG127. 2011. <http://guidance.nice.org.uk/CG127>.
- 43 Hermida RC, Ayala DE. Diagnosing gestational hypertension and pre-eclampsia with the 24-hour mean of blood pressure. *Hypertension* 1997;30:1531-7.
- 44 Hermida RC, Ayala DE. Reference thresholds for 24-h, diurnal, and nocturnal ambulatory blood pressure mean values in pregnancy. *Blood Pressure Monit* 2005;10:33-41.
- 45 Seventh report of the Joint National Committee on Prevention, Detection, Evaluation, and Treatment of High Blood Pressure. 2004. [www.nhlbi.nih.gov/guidelines/hypertension/jnc7full.htm](http://www.nhlbi.nih.gov/guidelines/hypertension/jnc7full.htm).
- 46 Lindheimer MD, Akbari A. Hypertension in pregnant women. In: Oparil S, Weber MA, eds. Hypertension: a companion to Brenner and Rector's the kidney. WB Saunders, 2000:688-701.
- 47 National Institutes of Health. Working group report on high blood pressure in pregnancy. National High Blood Pressure Education Program, 2000.
- 48 National Collaborating Centre for Women's and Children's Health. Antenatal care: routine care for the healthy pregnant woman. RCOG, 2008.
- 49 Ross-McGill H, Hewison J, Hirst J, Dowswell T, Holt A, Brunskill P, et al. Antenatal home blood pressure monitoring: a pilot randomised controlled trial. *Br J Obstet Gynaecol* 2000;107:217-21.
- 50 Thornton JG. Authors' reply. *BJOG* 2000;107:1181.
- 51 Aberg A. Diagnostic methods for pregnancy hypertension. *Int J Technol Assess Health Care* 1992;8(suppl 1):72-4.
- 52 Waugh J, Habiba MA, Bosio P, Boyce T, Shennan A, Halligan AW. Patient initiated home blood pressure recordings are accurate in hypertensive pregnant women. *Hypertens Pregnancy* 2003;22:93-7.

Accepted: 20 October 2014

Cite this as: *BMJ* 2014;349:g6616

© BMJ Publishing Group Ltd 2014