

Alignment of Klaus in Mount Wilson

Hale, Steven J.

License:

Creative Commons: Attribution-NonCommercial-ShareAlike (CC BY-NC-SA)

Document Version

Publisher's PDF, also known as Version of record

Citation for published version (Harvard):

Hale, SJ 2008 'Alignment of Klaus in Mount Wilson' BiSON Technical Report Series, no. 309, Birmingham Solar Oscillations Network. <<http://epapers.bham.ac.uk/2039/>>

[Link to publication on Research at Birmingham portal](#)

General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact UBIRA@lists.bham.ac.uk providing details and we will remove access to the work immediately and investigate.

BiSON Birmingham Solar-Oscillations Network

TECHNICAL REPORT NO. 309

Alignment of Klaus in Mount Wilson

Steven J. Hale

The University of Birmingham, Edgbaston, Birmingham B15 2TT

2008 August 1

This technical report series is published by:



**THE UNIVERSITY
OF BIRMINGHAM**

High-Resolution Optical-Spectroscopy Group

School of Physics and Astronomy
The University of Birmingham
Edgbaston, Birmingham B15 2TT, United Kingdom
Telephone: +44-121-414-4551 FAX: +44-121-414-1438

Alignment of Klaus in Mount Wilson

Steven J. Hale

The University of Birmingham, Edgbaston, Birmingham B15 2TT

2008 August 1

Abstract

The Pockels cell was emptied and refilled, all optics cleaned and aligned, a failing hard disk drive replaced, and the GPS repaired.

Contents

1	Introduction	1
2	Clean and Align All Optics	2
3	Replace Faulty Hard Disk	3
4	Refill the Pockels Cell	3
5	Fix the GPS	3
6	UPS	4

1 Introduction

Steven Hale visited Mount Wilson between 2008 May 12 and May 19. The main tasks that were planned for this trip were:

- Clean and align all optics.
- Replace a faulty hard disk.
- Refill the Pockels Cell.
- Fix the GPS.

2 Clean and Align All Optics

There are five mirrors in the system in front of Klaus.

The first two are the *coelostat* mirrors. They direct the beam vertically down through the sixty-foot tower. It was not possible to clean these two mirrors on this trip due to the lack of a nitrogen cylinder required to dry the surface after cleaning. This will be done at a later date.

About twenty feet from the bottom of the tower, there are two “periscope” mirrors, oriented at 45° to the vertical. They pick off a small part of the beam. The first periscope mirror deflects the beam sideways slightly, and the second deflects it down again.

At the bottom of the tower, the beam hits the fifth mirror and is reflected horizontally into Klaus. The final beam alignment is done with this fifth mirror. It is on an altitude/azimuth mount. Two micrometers allow it to be adjusted in the horizontal (azimuth) and the vertical (altitude) direction.

On this trip, the upwards-facing periscope mirror was removed and cleaned. The downwards-facing mirror was not very dirty. When reinstalling the mirror, adjustments were made to the periscope to ensure good alignment with the lower fifth mirror. Both the fifth mirror and the red filter at the front of the instrument were removed and cleaned.

There is a quadrant photodiode at the back of Klaus, and it was intended that this be used as an alignment monitor. Unfortunately the alignment monitor does not work — it doesn’t handle the photodiode signals properly and so it can’t actually be used to check the alignment.

There are two methods for checking how well the spectrometer is aligned with the fifth mirror. The first is by visually observing the path of the light within the instrument. For this to work the interference filter has to be removed in order to be able to see the beam. It is then possible to compare the position of the beam with the center of the lenses within the spectrometer, and also check for vignetting on the quadrant at the back of the instrument. The second method is more quantitative and usually more precise. This involves gradually stepping the mirror using the micrometers, one axis at a time, and plotting the effects on the data. It is then possible to determine from the graph the micrometer values that give the best data quality.

Both of these were done on a previous trip [1], with the visual method used to achieve basic alignment, and then the mirror scan to fine tune. However, the data quality were still not very good. The intensity profile was poor. The sum would rise during the morning and fall during the afternoon, rather than being flat throughout the day as with our other sites.

It seems that the best position indicated by the mirror scan results in some vignetting when observing the beam at the back of the instrument. Therefore on this trip spectrometer alignment was done using purely the visual method to ensure no vignetting occurred within the instrument, and no mirror scan was performed. The resulting data look much better. The sum is not totally flat, but it is greatly improved. The afternoon data are worse than the morning, but this is expected due to the additional haze caused by looking out over Los Angeles in the afternoon.

All cleaning and alignment was completed by 2230 on 2008 May 15.

The cleaning resulted in an increase in intensity of a few thousand counts. At our other sites such cleaning usually results in a jump in intensity of a significant percentage of the total signal. However in Mount Wilson the instrument is safely hidden away and so does not get as dirty — the *coelostat* mirrors take the brunt of the weather.

3 Replace Faulty Hard Disk

The master hard disk on the secondary IDE controller in the PC was showing signs of imminent failure. The monthly SMART scan would show up bad sectors on the disk. As a preventative measure the disk was replaced before it failed completely.

4 Refill the Pockels Cell

Some of the index matching fluid in the Pockel's cell had gone missing. There were no signs of the fluid having leaked, so where it goes remains a mystery. The cell was topped up, but the new fluid did not mix with the old and remained in two separate layers.

This meant that the cell had to be removed from the instrument by disconnecting the two SHVs in the connections box at the back, and all the fluid drained. The easiest way was to unscrew the front cover of the cell, marked "Do Not Open". With all the old liquid removed the cell was reassembled, refilled using just the new liquid, and reinstalled in the instrument.

The fluid used to fill the cell is known as Immersion Oil and supplied by Sigma-Aldrich. The liquid specification is shown in table 1.

Table 1: Immersion oil properties.

Product Number	51786-250ML
Product Line	BioChemika
Refractive Index	n ₂₀ /D 1.517
Viscosity	30-60 mPa.s (20°C)
Density	1.060 g/mL at 20°C
Hazard Codes	Xi
Risk Statements	36/37/38
Safety Statements	26

5 Fix the GPS

The GPS unit in Mt. Wilson has not worked for some time.

The cable from the GPS was traced outside towards the antenna. The fault was obvious. The antenna was no longer actually connected to the cable because the crimp on the BNC had failed.

A short section was cut from the antenna cable to remove some weather damage, and a new BNC crimped onto the end. The cable has now been cable tied to the building to take any strain off the connections, so hopefully the problem should not occur again.

The GPS started to work again as soon as the antenna was reconnected.

6 UPS

Shortly after a previous trip [1], Perry Rose bought a UPS for the main instrument in the sixty-foot tower. Since it had more than enough capacity, all of our electronics were moved over onto the UPS as well.

At some time in between then and now, the UPS failed. It was not replaced. So now none of our hardware is UPS protected any more.

References

- [1] STEVEN J. HALE. Alignment of Klaus in Mount Wilson. *BISON Technical Report Series*, Number 292, High-Resolution Optical-Spectroscopy Group, Birmingham, United Kingdom, August 2007.